

Spatial resolution limits for synchrotron-based infrared spectromicroscopy

Erika Levenson^a, Philippe Lerch^b, Michael C. Martin^{a,*}

^a Advanced Light Source Division, Lawrence Berkeley National Laboratory, 1 Cyclotron Road, Berkeley, CA 94720, USA

^b Swiss Light Source, Paul Scherrer Institut, 5232 Villigen, Switzerland

Available online 15 December 2007

Abstract

Detailed spatial resolution tests were performed on beamline 1.4.4 at the Advanced Light Source synchrotron facility in Berkeley, CA. The high-brightness synchrotron source is coupled at this beamline to a Thermo-Electron Continu μ m XL infrared microscope. Two types of resolution tests in both the mid-IR (using a KBr beamsplitter and an MCT-A* detector) and in the near-IR (using a CaF₂ beamsplitter and an InGaAS detector) were performed and compared to a simple diffraction-limited spot size model. At the shorter wavelengths in the near-IR the experimental results begin to deviate from only diffraction-limited. The entire data set is fit using a combined diffraction-limit and demagnified electron beam source size model. This description experimentally verifies how the physical electron beam size of the synchrotron source demagnified to the sample stage on the endstation begins to dominate the focussed spot size and therefore spatial resolution at higher energies. We discuss how different facilities, beamlines, and microscopes will affect the achievable spatial resolution. © 2007 Elsevier B.V. All rights reserved.

Keywords: Synchrotron; Spatial resolution; Microscopy; Spectromicroscopy; Imaging; Diffraction

1. Introduction

Synchrotron infrared (IR) beamlines provide diffraction-limited spatial resolution for spectromicroscopy with 100–1000 times higher brightness than a conventional thermal global source [1–4] enabling a wide variety of new science at small spatial scales [3,4]. We have previously experimentally verified the diffraction-limited performance through the mid-IR at the Advanced Light Source (ALS) [5]. However, as the synchrotron emission wavelength becomes shorter, the physical size of the electron beam will dominate the focused photon beam size.

The source size of a synchrotron light beam can be approximated well by adding in quadrature the effects of diffraction, the electron beam size, and the projected size of the emitting region [1,6]. This beam is then imaged onto a sample via beamline optics that have an overall demagnification factor, m . The spot size can therefore be written as

$$m\sqrt{(d\lambda)^2 + \sigma_t^2} \quad (1)$$

where d is a diffraction-limit factor (depends on which resolution model is chosen, as described later), λ is the wavelength of light, and σ_t is the transverse synchrotron electron beam size. The transverse electron beam size is given by

$$\sigma_t = \sqrt{\beta_t \varepsilon_t + \eta_t^2 \left(\frac{\sigma_E}{E}\right)^2} \quad (2)$$

where β_t is the beta-function, ε_t is the emittance, η_t is the dispersion (all in the transverse direction; t can be x or y), and σ_E/E is the energy spread [7]. These parameters are specific to each synchrotron light source and the specific beamline photon port.

2. Experiment

We performed lateral resolution experiments as a function of wavelength in the mid- and near-IR using a Thermo-Electron Continu μ m XL microscope and Nexus 870 FT-IR bench at ALS beamline 1.4.4. The beamline

* Corresponding author. Tel.: +1 510 495 2231; fax: +1 510 495 2067.
E-mail address: MCMartin@lbl.gov (M.C. Martin).

collects 10 mrad vertical by 40 mrad horizontal from the bending magnet source. A pair of cylindrical mirrors are used to collimate this source before the light is steered into the emission port of the FT-IR bench. The IR microscope focuses the beam onto the sample using all reflective 15 \times or 32 \times cassegrain objectives with numerical apertures of 0.58 and 0.65, respectively. The results presented here were obtained with the 32 \times objective. The sample stage is an automated Prior Scientific H101 stage with step sizes as small as 0.1 μm . All measurements were done in reflection mode without any apertures in the light path, and the results presented here detail the y -direction cross-sections of the focused spot.

We previously published mid-IR resolution test results showing the resolution is indeed diffraction-limited [5], so for this study we concentrate on the near-IR. An MCT-A* detector was used for 2000–7000 cm^{-1} , and an InGaAs detector was utilized for 5000–11,000 cm^{-1} . A CaF_2 beamsplitter was used to cover this entire range.

Spatial resolution tests were performed using a high-resolution USAF 1951 3-Bar Resolving Test Chart [8] from Applied Image, Inc. (Rochester, NY). It has a chrome metal coating on a glass substrate with the resolution test structures patterned in a negative image up to a spatial frequency of 512 cycles/mm (smallest is group 9, element 3).

Two types of resolution tests were done as defined in more detail in Ref. [5]: a step edge (or knife edge) test [9], and a Rayleigh's criterion test [10]. These two definitions are well known from the literature, however it is important to note that they yield different numerical results since they are based on different definitions. The Rayleigh imaging criterion will give a higher resolution than the FWHM step edge analysis by a factor 0.617.

2.1. Step edge resolution test results

Profiles were measured for each analyzed wavelength, then the derivative of each line profile was fit to a Gaussian [9]. Fig. 1 shows the FWHM resolution versus wavelength for both detectors. A simple linear fit to the data yields a resolution for the synchrotron source of $(0.73 \pm 0.04)\lambda$.

This, however, underestimates the slope of the real data, and the data deviate from the fit at shorter wavelengths. An improved and more physically meaningful fit is obtained by using Eq. (1). In this case the best fit results in the diffraction portion of the data given by $(0.81 \pm 0.02)\lambda$, with a demagnified electron beam size of $0.71 \pm 0.11 \mu\text{m}$.

2.2. Imaging resolution test results

Imaging tests were done using the USAF chart and Rayleigh's criterion [10] for $\lambda = 6.5$ – $1.11 \mu\text{m}$. The measured resolutions of all wavelengths analyzed are presented in Fig. 1. A simple linear fit to the data yields a resolution of $(0.45 \pm 0.02)\lambda$.

Again, however, this simple linear fit does not adequately follow the data points particularly at shorter wavelengths, nor does it intercept the origin. The experimental




Fig. 1. Step edge and Rayleigh resolution results as a function of wavelength. Fits to the data to the effective source size discussed in the text are shown as solid lines with the diffraction and electron beam source size components of each fit shown with dashed lines. (For a colour figure, the reader is referred to the web version of this article.)

imaging data clearly show deviation from a simple diffraction-limited spot size at short wavelengths. An improved fit is found using Eq. (1) with the diffraction portion of the data given by $(0.47 \pm 0.01)\lambda$, and a demagnified electron beam size of $0.51 \pm 0.06 \mu\text{m}$.

3. Analysis

As noted above, the step edge and imaging tests use different resolution definitions differing by a factor of 0.617. Indeed the diffraction portion of the best fits to Eq. (1) are different by a factor of 0.59 ± 0.03 and the electron beam size portions of the fits are different by a factor of 0.72 ± 0.20 , both within the error bars of the ideal factor. We conclude that the two resolution tests give consistent results.

The electron beam source size for the 1.4 (22.6°) bending magnet port at the ALS is $\sigma_x = 65 \mu\text{m}$ and $\sigma_y = 52 \mu\text{m}$ [11]. These are 1σ values, and the FWHM beam size is 2.35 times larger. Thus, the electron beam size is 153 μm in the horizontal and 122 μm in the vertical dimension.

The magnification factor m for the 1.4 beamline is given by the ratio of the focal distance of the collimating mirror to the focal distance of the microscope objective focusing the light onto the sample. The beamline 1.4.4 optics image the vertical synchrotron emission to the y -direction on the sample stage which is the direction in which the spot size measurements are detailed above. The magnification factor for the 32 \times objective is $m = 171$, and thus we would predict that the FWHM electron beam size of 122 μm is imaged onto the sample stage to 0.71 μm , FWHM. Or using the Rayleigh criterion, the imaged beam size is 0.44 μm . These values are in good agreement with experimental data of the demagnified electron beam source size fits of $0.71 \pm 0.11 \mu\text{m}$ and $0.51 \pm 0.06 \mu\text{m}$, for the two resolution

definitions, respectively. For the 32 \times objective with NA = 0.65, we expect the diffraction-limited performance of this objective to be 0.47λ . This is in excellent agreement with the fit to our measured data of $(0.47 \pm 0.01)\lambda$.

3.1. Resolution of different synchrotron beamlines

The achievable transverse resolution of a given synchrotron beamline is a combination of the beamline optics (demagnification factor), the electron beam source size which is dependent on the specific machine and photon port parameters, and the final focusing objective in the IR microscope. Thermo-Electron's 15 \times and 32 \times objectives have relatively large NA's, whereas Bruker Optics' 15 \times and 36 \times objectives have longer working distances but smaller NA's of 0.4 and 0.5, respectively. Using such objectives, the effects of the electron beam source size will start at longer wavelengths and will limit the ultimate lateral spatial resolution achievable. If we were to use the Bruker 15 \times objective (NA = 0.4) on the ALS IR beamline with all other optics being the same, we should obtain a spot size given by 0.76λ for the diffraction-limit, and $0.88\text{ }\mu\text{m}$ imaged electron beam size. Examples of the predicted resolution as a function of wavelength using different NA objectives are plotted in the upper panel of Fig. 2.

If the ALS IR beamline was placed on the smallest source size port at the ALS the vertical electron beam size would be $\sigma_y = 15\text{ }\mu\text{m}$, or a factor of 3.5 smaller. This would have minimal effect to the resolution in most of the mid-IR, but the diffraction-limited performance would extend further into the near-IR and a significant improvement in resolution could be achieved for wavelengths shorter than about $2\text{ }\mu\text{m}$. If the same beamline was built at the NSLS VUV ring (a second generation synchrotron which has IR microscopy beamlines with similar IR microscopes), the vertical electron beam source size is $\sigma_y = 185\text{ }\mu\text{m}$ [12], or about 3.5 times larger than the ALS 1.4 port. The electron beam source size will play a more dominant role in the total effective resolution, yielding about $0.5\text{ }\mu\text{m}$ larger spot size at $\lambda = 5\text{ }\mu\text{m}$ (2000 cm^{-1}), extending to $1\text{ }\mu\text{m}$ larger spot size at $\lambda = 2\text{ }\mu\text{m}$ (5000 cm^{-1}) compared to ALS BL1.4. An IR microscopy beamline is being built at the new SOLEIL synchrotron (France) with a planned $\sigma_y = 24.9\text{ }\mu\text{m}$ [13], so assuming similar magnification optics to the ALS the diffraction-limited performance will continue to wavelengths shorter than $1\text{ }\mu\text{m}$ ($10,000\text{ cm}^{-1}$).

The source size at the IR port currently being commissioned at the Swiss Light Source (SLS) storage ring is expected to be $\sigma_x = 52\text{ }\mu\text{m}$ and $\sigma_y = 26\text{ }\mu\text{m}$. The magnification factor m at the SLS IR beamline is 86 (with a 0.58 NA objective) or 82 (with a 0.61 NA objective). This means that the $\sigma_y = 26\text{ }\mu\text{m}$ vertical electron beam size will be imaged to the sample with a FWHM of $0.7\text{ }\mu\text{m}$, or $0.43\text{ }\mu\text{m}$ using the Rayleigh definition. These values are very close to those measured at the ALS so we would predict very similar resolution capabilities assuming the use of a similar microscope objective. The predicted resolution from several synchrotron source sizes are shown in the lower panel of Fig. 2.




Fig. 2. Predicted resolutions for different numerical aperture objectives, and different electron beam source sizes at selected synchrotron light sources.

The under development NSLS-II storage ring [14] is planned to perform at close to the theoretical minimum emittance [7] possible for a storage ring. The source size in the bending magnet ports will be $\sigma_x = 44.2\text{ }\mu\text{m}$ and $\sigma_y = 15.7\text{ }\mu\text{m}$. This is almost a factor of two smaller vertical beam size than SOLEIL, and so the diffraction limited performance should extend to approximately $\lambda = 400\text{ nm}$, in the visible. Different focusing optics could extend this range even further towards the VUV.

Acknowledgements

We would like to thank Fernando Sannibale. This work and the Advanced Light Source are supported by the Director, Office of Science, Office of Basic Energy Sciences, Materials Sciences Division, of the US Department of Energy under Contract No. DE-AC03-76SF00098 at Lawrence Berkeley National Laboratory.

References

- [1] J.A. Reffner, P.A. Martoglio, G.P. Williams, Rev. Sci. Instrum. 66 (1995) 1298;
G.L. Carr, J.A. Reffner, G.P. Williams, Rev. Sci. Instrum. 66 (1995) 1490;
G.L. Carr, Rev. Sci. Instrum. 72 (2001) 1613.

- [2] Michael C. Martin, Wayne R. McKinney, *Proc. Mater. Res. Soc.* 524 (1998) 1;
Michael C. Martin, Wayne R. McKinney, *Ferroelectrics* 249 (1–2) (2001) 1–10.
- [3] Hoi-Ying N. Holman, Michael C. Martin, Wayne R. McKinney, *Spectr. Int. J.* 17 (2–3) (2003) 139–159;
Hoi-Ying N. Holman, Michael C. Martin, *Adv. Agron.* 90 (2006) 79–127.
- [4] Paul Dumas, Mark J. Tobin, *Spectr. Eur.* 15 (6) (2003) 17–23;
L.M. Miller, P. Dumas, *Biochim Biophys Acta.* 1758 (2006) 846.
- [5] Erika Levenson, Philippe Lerch, Michael C. Martin, *Infrared Phys. Technol.* 49 (2006) 45–52.
- [6] C.J. Hirschmugl, Ph.D. Thesis, Yale University, 1994.
- [7] K.-J. Kim, Characteristics of synchrotron radiation, in: *Physics of Particle Accelerators*, AIP Conference Proceedings, vol. 184, American Institute of Physics, New York, 1989.
- [8] MIL-STD-150A, Section 5.1.1.7, USAF 1951 3-Bar Resolving Power Target; also see <<http://www.efg2.com/Lab/ImageProcessing/Test-Targets/>>.
- [9] John C. Russ, *The Image Processing Handbook*, CRC Press, 2002, p. 320.
- [10] Lord Rayleigh, *Phil. Mag.* 8 (5) (1879) 261;
Hecht, E. *Optics*, Addison Wesley Longman, Inc., 1998, p. 416;
Max Born, Emil Wolf, *Principles of Optics*, Cambridge University Press, 1999, p. 371.
- [11] <<http://www.als.lbl.gov/als/techspecs/srparameters.html>>.
- [12] <http://www.nsls.bnl.gov/facility/accelerator/vuv/vuv_parameters.pdf>.
- [13] <<http://www.synchrotron-soleil.fr/portal/page/portal/SourceAccelateur/ParametresPointsSources>>.
- [14] <http://www.bnl.gov/nsls2/project/source_properties>.